

**LUDLUM MODEL 43-10-10
ALPHA-BETA SAMPLE COUNTER**

**Addendum to Model 43-10-1 Manual
February 2020**

**LUDLUM MODEL 43-10-10
ALPHA-BETA SAMPLE COUNTER**

**Addendum to Model 43-10-1 Manual
February 2020**



LUDLUM MEASUREMENTS, INC
501 OAK STREET, P.O. BOX 810
SWEETWATER, TEXAS 79556
325-235-5494, FAX: 325-235-4672

LUDLUM MEASUREMENTS, INC.



501 OAK ST. / P.O. BOX 810
SWEETWATER, TEXAS 79556
Phone: 325-235-5494 800-622-0828 (USA)
Fax: 325-235-4672 E-Mail: ludlum@ludlums.com
Website: <http://www.ludlums.com> FEIN 75-1085764

**DESIGNER AND MANUFACTURER****OF***Scientific and Industrial
Instruments*



**Addendum for the Model 43-10-10 Alpha/Beta Sample Counter
with iCam Filter
(to the Model 43-10-1 manual)
February 2020**

The Model 43-10-10 (part # 47-4114), which allows for counting of iCam air monitor filter cards, is very similar to the Model 43-10-1. Following are the differences:

1. The sample drawer (or tray) has a cutout that fits the iCam filter card. The dimensions of the iCam filter card are 5.1 x 8.8 x 0.064 cm (2 x 3.5 x 0.025 in.) (H x W x D). It can also fit a 60 x 4.7 mm deep (2.36 x 0.187 in.) sample.
2. Other new part numbers include:
 - Tube holder/Base assembly – part # 2142-303
 - Tray – part # 7142-320
 - Base plate – 7142-305
 - 60 mm spacer – 7142-329
3. See Drawing 142 x 301 (following this page) for overall dimensions.



REVISION HISTORY			
REV	DESCRIPTION	DATE	BY
1	VALID	12-16-14	CMC



DWN	DATE	CHK	DATE	APP	DATE
JGW	2-11-20			<i>lw</i>	2-11-20
DWG NUM: 4142-301				SCALE: 7/16	
TITLE M 43-10-10 OVERALL					
LUDLUM MEASUREMENTS, INC. 501 OAK STREET SWEETWATER, TEXAS 79556			SERIES 142	SHEET 301	

**Model 43-10-1 Alpha/Beta Sample Counter
December 2019**

TABLE OF CONTENTS

1. GENERAL-----	1
2. SPECIFICATIONS-----	1
3. OPERATING PROCEDURES-----	2
4. CALIBRATION-----	2
4.1 Counting Instrument-----	2
4.2 Operating Voltage-----	2
4.3 Calculating Efficiency-----	3
5. TROUBLESHOOTING-----	3
5.1 Zero or Very Low Counts-----	3
5.2 No Source Plateau-----	3
5.3 Excessive Background Count-----	3
PARTS LIST-----	4
Model 43-10-1 Alpha/Beta Sample Counter-----	4
Assembly View, Drawing 142 X 39B-----	4
Switch Filter Board, Drawing 142 X 58-----	4
Voltage Divider Board Board, Drawing 435 X 964-----	4
DRAWINGS AND DIAGRAMS-----	5

Model 43-10-1 Alpha/Beta Sample Counter
December 2019

1. GENERAL

The Model 43-10-1 is an Alpha-Beta Sample Counter capable of holding up to a 5.1 cm (2 in.) diameter filter or planchet. The sample drawer, when fully closed, strikes a microswitch to allow high voltage (HV) to be applied to the photomultiplier tube (PMT). The sample drawer is locked in the closed position by rotation of the slide lever mounted on the side of the instrument.

The detector is a 6.4 cm (2.5 in.) diameter "phoswich" with a 0.025 cm (0.010 in.) thick plastic scintillator coated with zinc sulfide (ZnS).

ZnS(Ag) is used for alpha radiation detection, and the plastic scintillation material is used for detection of beta radiation. The scintillation material is covered by 0.4 mg/cm² metalized polyester to reduce light response (excessive background). If simultaneous alpha-beta discrimination is desired, the counting instrument must have separate power supplies or threshold controls for each channel. The Ludlum Model 2929 Scaler, Model 2223, or Model 2224 instruments provide the necessary circuitry for simultaneous alpha-beta discrimination.

2. SPECIFICATIONS

SCINTILLATOR MATERIAL: ZnS disc; plastic 0.025 cm (0.010 in.) thick

DETECTOR OPERATING VOLTAGE: 500-1200 Vdc

WINDOW: 0.4 mg/cm²

BACKGROUND:
≤ 80 cpm beta-gamma, ≤ 3 cpm alpha
(in ambient background of 10μR/hr)

CHANNEL CROSS TALK: alpha in beta channel ≤ 10%; beta in alpha channel ≤ 1%

EFFICIENCY (4π): 37% for ²³⁹Pu, 5% for ¹⁴C, 27% for ⁹⁹Tc, 32% for ²³⁰Th, 39% for ²³⁸U, 29% for ¹³⁷Cs, 26% for ⁹⁹Sr/⁹⁰Y

HV SWITCH: opening sample slide disables PMT high voltage

PHOTOMULTIPLIER TUBE: 5.1 cm (2 in.) diameter, 10 pin dynode structure

SAMPLE SLIDE AND HOLDER: sample cavity size is 56.9 mm (2.24 in.) diameter x 10.8 mm (0.428 in.) deep, with an insert cavity size of 50.8 mm (2.0 in.) diameter x 4.4 mm (0.175 in.) deep or 28.3 mm (1.115 in.) diameter x 4.4 mm (0.175 in.) deep.

MAXIMUM SAMPLE SIZE: 56.9 mm (2.24 in.) diameter x 10.8 mm (0.428 in.) deep

CONSTRUCTION: aluminum housing with beige powder coating

SIZE: 23.6 x 11.4 x 23.6 cm (9.3 x 4.5 x 9.3 in.) (H x W x L)

Model 43-10-1 Alpha/Beta Sample Counter
December 2019

3. OPERATING PROCEDURES

Connect the Model 43-10-1 to the scaler counting instrument. The coax cable with "C" connectors carries both the signal and HV.

HV is applied to the PMT when the sample slide is pushed completely in, tripping the microswitch. Rotate the sample slide lever to the locked position, securing sample slide in the "ON" position.

Alpha background count is approximately less than or equal to 3 cpm.

Beta background count is approximately 60-100 cpm.

To check a radioactive sample, place sample on the appropriate side of the sample holder for the 2.5 or 5.1-centimeter (1 or 2 in.) filters. Do not allow the sample to extend above the top of the sample slide.

A background count should be taken after each source count to check for contamination on the sample holder or area within the O-ring.

4. CALIBRATION

CAUTION: Do not tip sample counter over with sample holder in sample slide. The sample holder will tear the thin metalized polyester window, allowing light to scintillate the ZnS and cause excessive count in the beta channel.

For instruments with separate power supplies (fixed threshold), the alpha channel will operate at a lower voltage than the beta channel.

4.1 Counting Instrument

Calibrated scaler instrument
HV range, nominally 800 ±200 volts
Nominal input sensitivity:
alpha channel = 175 mV
beta channel = 4 mV (with upper discriminator set at 50 mV)

4.2 Operating Voltage

1. Connect Model 43-10-1 to the counting instrument with proper cable.
2. Place a calibrated ¹⁴C source in the

sample holder. Close and lock the sample drawer.

3. Adjust the counting instrument HV until it receives at least 5% (4π) efficiency.
4. Decrease HV by 25 volts.
5. Record the HV.
6. Record the ¹⁴C source count and beta crosstalk in the alpha channel.
7. Remove the ¹⁴C source and record the background count in the alpha and beta channels.
8. Place a calibrated ²³⁹Pu source in the sample holder. Close and lock the sample drawer.
9. Record the ²³⁹Pu source count and the alpha crosstalk in the beta channel.
10. Increase the HV by 25 volts.
11. Repeat steps 5-10 until one or more of the following conditions is met (assuming a 10 μR/hr background exposure):
 - (a) beta background exceeds 80 cpm

**Model 43-10-1 Alpha/Beta Sample Counter
December 2019**

- (b) alpha background exceeds 3 cpm
 - (c) alpha crosstalk in the beta channel exceeds 10%
 - (d) beta crosstalk in the alpha channel exceeds 1%
12. The operating voltage should be selected as a point where:
- (a) ^{14}C efficiency (4π) $\geq 5\%$
 - (b) ^{239}Pu efficiency (4π) $\geq 37\%$
 - (c) alpha crosstalk in beta channel less than or equal to 10%
 - (d) beta crosstalk in alpha channel less than or equal to 1%

4.3 Calculating Efficiency

1. NIST-traceable sources required.

2. Set HV as determined above.
3. Record a one-minute background and one-minute source count. Subtract the background count from the source count. Divide the net source count by the dpm value of the source, times 100 for 4π efficiency.

If the source value is listed in microcuries (activity):

4. Convert the microcurie value to a dpm value by multiplying the microcurie value by 2.22×10^6 . Calculate the 4π efficiency as in the previous steps.

5. TROUBLESHOOTING

5.1 Zero or Very Low Counts

- Large light leak
- PMT malfunction
- Broken wire in tube socket
- Inoperative HV switch on sample counter or broken wire
- Counting instrument malfunction
- Source too far from scintillation material
- Cable malfunction

5.2 No Source Plateau

- Light leak, slide not sealed properly against true base
- Noisy PMT
- Noisy HV switch
- Poor PMT to scintillation, light pipe interface

5.3 Excessive Background Count

- Light leak
- PMT malfunction
- Cable malfunction
- Noisy HV switch
- Instrument contaminated

**Model 43-10-1 Alpha/Beta Sample Counter
December 2019**

PARTS LIST

Ref. No.	Description	Part No.	Ref. No.	Description	Part No.
Model 43-10-1 Alpha/Beta Sample Counter			Switch Filter Board, Drawing 142 X 58		
UNIT	Completely Assembled 43-10-1 Detector	47-1305	BOARD	Assembled Switch Filter	5412-103
Assembly View, Drawing 142 x 39B			<ul style="list-style-type: none"> ▪ CAPACITORS C1-C2 CAP-0.0047μF, 3kV, NPO 04-5547 C3 CAP-0.0015μF, 3kV, C 04-5518 ▪ RESISTORS R1-R2 RES-1MEG, 1/4W, 5% 10-7028 		
* PM TUBE ASSY		01-5919	Voltage Divider Board, Drawing 435 X 964		
* EJ444L-2.20 x .010 ZnS		01-5698	BOARD	Assembled Voltage Divider	5435-401
* METALIZED MYLAR		01-5143	<ul style="list-style-type: none"> ▪ CAPACITORS C1 0.01μF, 2kV, C 04-5722 ▪ RESISTORS R1-R12 4.75 MEG, 1/8W, 1% 12-7995 		
* TUBE HOLDER/BASE		2142-002-02			
* CONNECTOR CAP		7142-014			
* SAMPLE DRAWER					
Model 43-10		7142-001-06			
* O-RING-2-229		16-8286			
* ACRYLIC DISC		7142-002-01			
2 EA. SPACER STRIP .015		7142-002-03			
* ADAPTER PLATE		7142-003-01			
* CASE TOP		7142-004-03			
* CASE BOTTOM		7142-004-04			
* CAP GASKET		7142-017			
* BASE PLATE		7142-018			
* SHAFT		7142-019			
* LIFTER		7142-020			
* PIN		7142-021			
* O-RING-2-226		16-8270			
2 EA. SPACER STRIP .010		7142-232			
5 EA. 5.1 cm (2 in.) X-TAL FOAM PAD					
		7260-001-05			
10 EA. PLANCHETTE-2/X1/8 IN.		7525-371A			
* PLANCHET HOLDER		7142-001-07			
* BRACKET		7142-004-01			
* CAP		7142-004-02			
1 EA. SWITCH-BZ-2RD-A2		08-6538			
1 EA. KNOB-90 4 2G POINTER		08-6608			
1 EA. RECPT-UG706/U "C" LMI		4478-011			
4 EA. BUMPER PADS		21-9376			
2 EA. SPACERS		18-9043			

DRAWINGS AND DIAGRAMS


Model 43-10-1 Assembly View, Drawing 142 x 39B

Switch Filter Board, Drawing 142 x 58

Switch Filter Board Layout, Drawing 142 x 59

5.1 cm (2 in.) Voltage Divider Board, Drawing 435 x 964

5.1 cm (2 in.) Voltage Divider Board Layout, Drawing 435 x 965




DETAIL L
SCALE 3.13:1

REVISION HISTORY


REV	DESCRIPTION	DATE	BY
3	ECF # 1072	10/31/01	DSW
4	REDRAWN ON COMPUTER	05/23/2012	ADG
5	PLACED LOCKWASHERS	1/2/18	ZSZ
6	17-8518 WAS 17-8511	11/19/19	DLJ


DWN	DATE	CHK	DATE	APP	DATE
ABM	11/27/19			Jew	11-27-19
DWG NUM: 4142-076				SCALE: 1:1.5	
TITLE M 43-10-1					
LUDLUM MEASUREMENTS, INC. 501 OAK STREET SWEETWATER, TEXAS 79556			SERIES	SHEET	
			142	39B	


REVISIONS						
EFF	AUTHORITY	ZONE	LTR	DESCRIPTION	DATE	APPROVED




UPDATED	-	LUDLUM MEASUREMENTS INC.			
DR PW	10/20/92	TITLE: SWITCH FILTER BOARD			
CHK	CKB 27-JAN-99				
DSCN PW	10/20/92	BOARD# 5142-103			
APPD	ESS 11-6-01	SIZE	MODEL	SERIES	SHEET
NEXT HIGHER ASSY.	-	C	43-10	142	58
08:36:53	27-Jan-99	58142103	SHEET 1 OF 1		



 LUDLUM MEASUREMENTS INC. SWEETWATER, TX.			
DR	PW	10/20/92	TITLE: SWITCH
CHK	CKR	27-JAN-99	FILTER BOARD
DSCN	PW	10/20/92	BOARD# S142-103
APP	CS	1-27-99	BS142103
07:28:59	27-Jan-99	MODEL 43-10	SERIES 142
COMP PASTE	COMP MASK	SLDR PASTE	SHEET 59
			S9
			OUTLINE



		PO Box 810 501 Oak Street Sweetwater, Texas 79556 U.S.A. 1-800-622-0828	
Drawn: AC	05/07/2012	Title: VOLTAGE DIVIDER	
Design: RSS	05/07/2012	Model: VARIOUS	
		Board#: 5435-401	
Approve: <i>J.W.</i>	<i>10-22-12</i>	Sheet: 1 of 1	Series
Print Date: 10/10/2012 9:37:50 AM	Rev: 2		Sheet
W:\Projects\LMI\VoltageDividers\5435-401\Rev2\435401R2P1.SchDoc		435	964



LUDLUM
MEASUREMENTS, INC.

PO Box 810
501 Oak Street
Sweetwater, TX 79556
U.S.A. 1-800-622-0828

Title: VOLTAGE DIVIDER				
Drawn: AC	05/07/2012	Model: VARIOUS		
Design: RSS	05/07/2012	Board#: 5435-401		
Approve: <i>AW</i>	<i>10-22-12</i>	Rev: 2		
Print Date:		SCALE: 1.00	Series	Sheet
10/10/2012	9:37:52 AM	Top Overlay	435	965
W:\Projects\LMIVoltageDividers\5435-401\Rev2\435401R2_Manual.PcbDoc				