

MET LABORATORIES, INC. CERTIFICATION RECORD



The applicant named below has been authorized by MET Laboratories, Inc. to represent the product(s) listed in this record as "MET Certified" and to mark this/these product(s) according to the terms and conditions of the MET Mark Utilization Agreement, MET Listing Reports, and the applicable marking agreements. Only the product(s) bearing the MET Mark and under a follow-up service are considered to be included in the MET Certification program.

FILE NUMBER: E112614

APPROVAL DATE: August 3, 2007

REVISED: November 12, 2008

PRODUCT	MODEL	ELECTRICAL RATINGS
Survey Meter	Models 3 and 3-IS-1	3Vdc - provided by two D-cell, alkaline batteries (Duracell Procell, model PC1300) Class I, Division 1, Groups A, B, C, and D, T6

STANDARD NUMBER	STANDARD TITLE	EDITION
UL61010-1 & CAN/CSA-C22.2 No. 61010-1	Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use – Part 1: General Requirements	Second
UL 913	Intrinsically Safe Apparatus and Associated Apparatus for Use in Class I, II, and III, Division 1, Hazardous (Classified) Locations	Sixth
CAN/CSA-C22.2 No. 157-92 (R2002)	Intrinsically Safe and Non-incendive Equipment for Use in Hazardous Locations	Third

MET LABORATORIES, INC. requires that any and all changes proposed in the previously identified product(s), that affects the information contained in the above referenced listing report, must be submitted to MET for evaluation prior to implementation to assure continued MET Certification status.

The above identified product(s) has/have been submitted by the applicant:

APPLICANT:

Ludlum Measurements, Inc.
501 Oak Street
Sweetwater, Texas 79556 U.S.A.

The covered products shall be subjected to follow-up inspections to ensure that the Certified product(s) are identical to the representative product sample evaluated by MET LABORATORIES, INC. and that all manufacturer's responsibilities are being fulfilled as specified in the MANUFACTURING RESPONSIBILITY section of the Certification report.

Rick Cooper
Director of Laboratory Operations,
Safety Laboratory



*MET Laboratories, Inc. is accredited by OSHA and the Standards Council of Canada.
The Nation's First Nationally Recognized Testing Laboratory*

NRTL